

SOLUTION CHANGE

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 01/06/2021

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:46
INTERNAL STANDARD	VERIFIED	20:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	20:47
BLANK TEST	0.00	20:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:48
BLANK TEST	0.00	20:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:49
BLANK TEST	0.00	20:50

