

SUPERVISOR TEST

-----  
HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408  
-----

Date: 01/12/2021

OPERATOR NAME:  
RICHARD J ARNOLD  
Permit #: 33860  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:56
INTERNAL STANDARD	VERIFIED	19:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:57
BLANK TEST	0.00	19:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:58
BLANK TEST	0.00	19:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:59
BLANK TEST	0.00	20:00

