

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 01/20/2021

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:45
INTERNAL STANDARD	VERIFIED	18:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:45
BLANK TEST	0.00	18:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:47
BLANK TEST	0.00	18:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:48
BLANK TEST	0.00	18:49

