

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 01/11/2021

OPERATOR NAME:
CARLEIGH B BATCHELDER
Permit #: 48533
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:34
INTERNAL STANDARD	VERIFIED	23:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:34
BLANK TEST	0.00	23:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:36
BLANK TEST	0.00	23:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:37
BLANK TEST	0.00	23:38

CB # 708

