

SUPERVISOR TEST

-----  
 CAMBRIDGE VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 126706  
 -----

Date: 01/19/2021

OPERATOR NAME:  
 ROBERT M DANKO  
 Permit #: 39105  
 Operator Agency/Dept: CAMBRIDGE VILL. PD  
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:01
INTERNAL STANDARD	VERIFIED	15:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:02
BLANK TEST	0.00	15:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:03
BLANK TEST	0.00	15:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:04
BLANK TEST	0.00	15:05

