

SOLUTION CHANGE

 CAMBRIDGE VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 126706

Date: 12/30/2020

OPERATOR NAME:
 ROBERT M DANKO
 Permit #: 39105
 Operator Agency/Dept: CAMBRIDGE VILL. PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:14
INTERNAL STANDARD	VERIFIED	15:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:14
BLANK TEST	0.00	15:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:15
BLANK TEST	0.00	15:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:17
BLANK TEST	0.00	15:17

