

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 01/24/2021

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:37
INTERNAL STANDARD	VERIFIED	21:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:38
BLANK TEST	0.00	21:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:39
BLANK TEST	0.00	21:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:40
BLANK TEST	0.00	21:41

