

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 02/15/2021

OPERATOR NAME:  
ROBERT J SULLIVAN  
Permit #: 34106  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:08
INTERNAL STANDARD	VERIFIED	00:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:09
BLANK TEST	0.00	00:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:10
BLANK TEST	0.00	00:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:11
BLANK TEST	0.00	00:12

