

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 02/17/2021

OPERATOR NAME:
 RYAN M WOODWARD
 Permit #: 41955
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:36
INTERNAL STANDARD	VERIFIED	13:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:37
BLANK TEST	0.00	13:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:38
BLANK TEST	0.00	13:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:39
BLANK TEST	0.00	13:40

