

SUPERVISOR TEST

WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 02/15/2021

OPERATOR NAME:
 CHRISTOPHER J MURRAY
 Permit #: 46226
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:35
INTERNAL STANDARD	VERIFIED	00:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:36
BLANK TEST	0.00	00:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:37
BLANK TEST	0.00	00:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:38
BLANK TEST	0.00	00:39

CM #821

