

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 02/28/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:43
INTERNAL STANDARD	VERIFIED	05:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:44
BLANK TEST	0.00	05:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:45
BLANK TEST	0.00	05:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:46
BLANK TEST	0.00	05:47

