

SUPERVISOR TEST

-----  
WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284  
-----

Date: 03/08/2021

OPERATOR NAME:  
JOSHUA T WHITNEY  
Permit #: 40382  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:43
INTERNAL STANDARD	VERIFIED	15:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:44
BLANK TEST	0.00	15:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:45
BLANK TEST	0.00	15:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:46
BLANK TEST	0.00	15:47

