

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 03/15/2021

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:33
INTERNAL STANDARD	VERIFIED	06:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:33
BLANK TEST	0.00	06:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:34
BLANK TEST	0.00	06:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:36
BLANK TEST	0.00	06:36

