

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 04/04/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:54
INTERNAL STANDARD	VERIFIED	02:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:55
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:56
BLANK TEST	0.00	02:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:57
BLANK TEST	0.00	02:58

