

SOLUTION CHANGE

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 04/01/2021

OPERATOR NAME:
 ERIC C WRIGHT
 Permit #: 49047
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 1104

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:09
INTERNAL STANDARD	VERIFIED	15:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:10
BLANK TEST	0.00	15:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:11
BLANK TEST	0.00	15:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:12
BLANK TEST	0.00	15:13

