

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 03/22/2021

OPERATOR NAME:
KATHERINE M PADDOCK
Permit #: 45899
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:00
INTERNAL STANDARD	VERIFIED	15:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:00
BLANK TEST	0.00	15:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:01
BLANK TEST	0.00	15:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:03
BLANK TEST	0.00	15:04

Handwritten signature/initials

