

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 04/21/2021

OPERATOR NAME:
BRAD M LYON
Permit #: 44978
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:35
INTERNAL STANDARD	VERIFIED	10:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:36
BLANK TEST	0.00	10:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:37
BLANK TEST	0.00	10:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:38
BLANK TEST	0.00	10:39

