

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 04/12/2021

OPERATOR NAME:  
CARLEIGH B BATCHELDER  
Permit #: 48533  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:44
INTERNAL STANDARD	VERIFIED	23:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:44
BLANK TEST	0.00	23:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:46
BLANK TEST	0.00	23:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:47
BLANK TEST	0.00	23:48

~~WJG~~ <sup>WJS</sup>  
~~647~~

