

SOLUTION CHANGE

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 05/01/2021

OPERATOR NAME:
 ROBERT J SULLIVAN
 Permit #: 34106
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:29
INTERNAL STANDARD	VERIFIED	09:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:29
BLANK TEST	0.00	09:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:30
BLANK TEST	0.00	09:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:31
BLANK TEST	0.00	09:32

