

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 05/12/2021

OPERATOR NAME:
RYAN M WOODWARD
Permit #: 41955
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:16
INTERNAL STANDARD	VERIFIED	14:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:16
BLANK TEST	0.00	14:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:17
BLANK TEST	0.00	14:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:19
BLANK TEST	0.00	14:20

