

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 05/31/2021

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:26
INTERNAL STANDARD	VERIFIED	16:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:27
BLANK TEST	0.00	16:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:28
BLANK TEST	0.00	16:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:29
BLANK TEST	0.00	16:30

