

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 05/23/2021

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:04
INTERNAL STANDARD	VERIFIED	23:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:04
BLANK TEST	0.00	23:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:05
BLANK TEST	0.00	23:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:07
BLANK TEST	0.00	23:08

