

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 06/14/2021

OPERATOR NAME:
CHRISTOPHER J DAVIDSEN
Permit #: 40077
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:46
INTERNAL STANDARD	VERIFIED	16:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:47
BLANK TEST	0.00	16:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:48
BLANK TEST	0.00	16:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:49
BLANK TEST	0.00	16:50

