

UPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 07/05/2021

OPERATOR NAME:  
CORY L HURLBURT  
Permit #: 37017  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:15
INTERNAL STANDARD	VERIFIED	09:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:16
BLANK TEST	0.00	09:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:17
BLANK TEST	0.00	09:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:18
BLANK TEST	0.00	09:19

#257

