

SOLUTION CHANGE

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 07/01/2021

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:37
INTERNAL STANDARD	VERIFIED	07:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:37
BLANK TEST	0.00	07:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:38
BLANK TEST	0.00	07:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:40
BLANK TEST	0.00	07:40

