

SUPERVISOR TEST

-----  
 WASHINGTON CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 100284  
 -----

Date: 07/06/2021

OPERATOR NAME:  
 MATTHEW K JACKSON  
 Permit #: 40078  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:22
INTERNAL STANDARD	VERIFIED	06:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:23
BLANK TEST	0.00	06:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:24
BLANK TEST	0.00	06:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:25
BLANK TEST	0.00	06:26

