

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 07/12/2021

OPERATOR NAME:
CHRISTOPHER J DAVIDSEN
Permit #: 40077
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:51
INTERNAL STANDARD	VERIFIED	08:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:51
BLANK TEST	0.00	08:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:52
BLANK TEST	0.00	08:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:54
BLANK TEST	0.00	08:54

