

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 07/17/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:52
INTERNAL STANDARD	VERIFIED	22:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:52
BLANK TEST	0.00	22:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:53
BLANK TEST	0.00	22:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:55
BLANK TEST	0.00	22:55

