

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 07/19/2021

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:07
INTERNAL STANDARD	VERIFIED	07:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:07
BLANK TEST	0.00	07:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:08
BLANK TEST	0.00	07:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:09
BLANK TEST	0.00	07:10

