

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 07/26/2021

OPERATOR NAME:
ERIC C WRIGHT
Permit #: 49047
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21100

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 16:53 |
| INTERNAL STANDARD | VERIFIED | 16:53 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 16:54 |
| BLANK TEST | 0.00 | 16:55 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 16:55 |
| BLANK TEST | 0.00 | 16:56 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 16:56 |
| BLANK TEST | 0.00 | 16:57 |

