

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 08/10/2021

OPERATOR NAME:
JOSEPH F MCKEEVER
Permit #: 49488
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:38
INTERNAL STANDARD	VERIFIED	22:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:38
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:40
BLANK TEST	0.00	22:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:41
BLANK TEST	0.00	22:42

