

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 08/16/2021

OPERATOR NAME:
DAVID M COSTELLO
Permit #: 40891
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:39
INTERNAL STANDARD	VERIFIED	23:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:39
BLANK TEST	0.00	23:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:40
BLANK TEST	0.00	23:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:41
BLANK TEST	0.00	23:42

