

SOLUTION CHANGE

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 08/09/2021

OPERATOR NAME:  
DAVID F BUXTON  
Permit #: 33398  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:06
INTERNAL STANDARD	VERIFIED	16:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:06
BLANK TEST	0.00	16:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:07
BLANK TEST	0.00	16:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:08
BLANK TEST	0.00	16:09

