

SOLUTION CHANGE

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 08/03/2021

OPERATOR NAME:
 SGT DEAN E WATKINS
 Permit #: 39523
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:51
INTERNAL STANDARD	VERIFIED	00:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:52
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:53
BLANK TEST	0.00	00:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:54
BLANK TEST	0.00	00:55

