

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 08/28/2021

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:57
INTERNAL STANDARD	VERIFIED	19:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:58
BLANK TEST	0.00	19:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:59
BLANK TEST	0.00	20:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:00
BLANK TEST	0.00	20:01

