

SUPERVISOR TEST

-----  
GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306  
-----

Date: 08/17/2021

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:13
INTERNAL STANDARD	VERIFIED	10:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:13
BLANK TEST	0.00	10:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:14
BLANK TEST	0.00	10:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:16
BLANK TEST	0.00	10:16

