

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 09/13/2021

OPERATOR NAME:
D TODD RICHE
Permit #: 33411
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:24
INTERNAL STANDARD	VERIFIED	08:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.08	08:24
BLANK TEST	0.00	08:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.08	08:25
BLANK TEST	0.00	08:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.08	08:26
BLANK TEST	0.00	08:27

