

SUPERVISOR TEST

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 08/30/2021

OPERATOR NAME:
 SGT DEAN E WATKINS
 Permit #: 39523
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:38
INTERNAL STANDARD	VERIFIED	22:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:39
BLANK TEST	0.00	22:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:40
BLANK TEST	0.00	22:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:42
BLANK TEST	0.00	22:42

