

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 09/27/2021

OPERATOR NAME:
JEFFREY E PETERMAN
Permit #: 49748
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 14:43 |
| INTERNAL STANDARD | VERIFIED | 14:43 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 14:43 |
| BLANK TEST | 0.00 | 14:44 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 14:44 |
| BLANK TEST | 0.00 | 14:45 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 14:46 |
| BLANK TEST | 0.00 | 14:47 |

