

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 09/22/2021

OPERATOR NAME:  
DAVID M COSTELLO  
Permit #: 40891  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:20
INTERNAL STANDARD	VERIFIED	23:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:20
BLANK TEST	0.00	23:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:21
BLANK TEST	0.00	23:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:22
BLANK TEST	0.00	23:23

