

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 10/18/2021

OPERATOR NAME:
 D TODD RICHE
 Permit #: 33411
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:08
INTERNAL STANDARD	VERIFIED	08:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:08
BLANK TEST	0.00	08:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:09
BLANK TEST	0.00	08:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:10
BLANK TEST	0.00	08:11

