

SUPERVISOR TEST

-----  
 WASHINGTON COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106406  
 -----

Date: 10/25/2021

OPERATOR NAME:  
 WILLIAM M GREENE  
 Permit #: 48760  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:30
INTERNAL STANDARD	VERIFIED	14:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:30
BLANK TEST	0.00	14:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:31
BLANK TEST	0.00	14:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:33
BLANK TEST	0.00	14:34

