

SOLUTION CHANGE

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 11/01/2021

OPERATOR NAME:
DAVID M COSTELLO
Permit #: 40891
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:28
INTERNAL STANDARD	VERIFIED	00:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:29
BLANK TEST	0.00	00:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:30
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:31
BLANK TEST	0.00	00:32

