

SUPERVISOR TEST

HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408

Date: 11/11/2021

OPERATOR NAME:  
 DAVID M COSTELLO  
 Permit #: 40891  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:33
INTERNAL STANDARD	VERIFIED	10:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:34
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:35
BLANK TEST	0.00	10:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:36
BLANK TEST	0.00	10:37

