

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 11/15/2021

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:44
INTERNAL STANDARD	VERIFIED	06:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:45
BLANK TEST	0.00	06:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:46
BLANK TEST	0.00	06:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:47
BLANK TEST	0.00	06:48

