

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 11/24/2021

OPERATOR NAME:
 LANCE M BARBER
 Permit #: 42841
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:07
INTERNAL STANDARD	VERIFIED	19:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:08
BLANK TEST	0.00	19:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:09
BLANK TEST	0.00	19:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:10
BLANK TEST	0.00	19:11

