

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 12/04/2021

OPERATOR NAME:
ALEXANDER T GILMORE
Permit #: 47649
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:39
INTERNAL STANDARD	VERIFIED	23:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:40
BLANK TEST	0.00	23:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:41
BLANK TEST	0.00	23:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:42
BLANK TEST	0.00	23:43

