

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 12/20/2021

OPERATOR NAME:
TREVOR J PECK
Permit #: 49747
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:16
INTERNAL STANDARD	VERIFIED	01:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:17
BLANK TEST	0.00	01:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:18
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:19
BLANK TEST	0.00	01:20

12/20
037
77P

