

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 12/02/2021

OPERATOR NAME:
ERNEST R BASSETT
Permit #: 25518
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:55
INTERNAL STANDARD	VERIFIED	21:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:55
BLANK TEST	0.00	21:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:56
BLANK TEST	0.00	21:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:58
BLANK TEST	0.00	21:58

