

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 12/28/2021

OPERATOR NAME:  
ERNEST R BASSETT  
Permit #: 25518  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:13
INTERNAL STANDARD	VERIFIED	06:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:14
BLANK TEST	0.00	06:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:15
BLANK TEST	0.00	06:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:16
BLANK TEST	0.00	06:17

